



GAS EQUIPMENT TECHNOLOGY GROUP

UNITS : SLPM NITROUS OXIDE DATE : 07-09-96
TUBE NUMBER : #6 (E606) STD CONDITIONS : 1 ATMOS. & 70 DEG F
SERIAL NUMBER : 10124 GAS TEMPERATURE : 70 DEG F
FLOAT MATERIAL: 316 ST. STL. PRESSURE IN TUBE : 0 PSIG

SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE
150.0	6.66	110.0	5.11	70.0	3.46	30.0	1.74
149.0	6.62	109.0	5.07	69.0	3.42	29.0	1.70
148.0	6.58	108.0	5.03	68.0	3.38	28.0	1.65
147.0	6.55	107.0	4.99	67.0	3.33	27.0	1.61
146.0	6.51	106.0	4.95	66.0	3.29	26.0	1.56
145.0	6.47	105.0	4.91	65.0	3.25	25.0	1.52
144.0	6.44	104.0	4.87	64.0	3.21	24.0	1.48
143.0	6.40	103.0	4.83	63.0	3.16	23.0	1.43
142.0	6.36	102.0	4.79	62.0	3.12	22.0	1.39
141.0	6.32	101.0	4.75	61.0	3.08	21.0	1.34
140.0	6.29	100.0	4.71	60.0	3.04	20.0	1.30
139.0	6.25	99.0	4.67	59.0	2.99	19.0	1.26
138.0	6.21	98.0	4.63	58.0	2.95	18.0	1.21
137.0	6.17	97.0	4.58	57.0	2.91	17.0	1.17
136.0	6.13	96.0	4.54	56.0	2.86	16.0	1.12
135.0	6.09	95.0	4.50	55.0	2.82	15.0	1.08
134.0	6.06	94.0	4.46	54.0	2.78	14.0	1.03
133.0	6.02	93.0	4.42	53.0	2.74	13.0	0.99
132.0	5.98	92.0	4.38	52.0	2.69	12.0	0.94
131.0	5.94	91.0	4.34	51.0	2.65	11.0	0.90
130.0	5.90	90.0	4.30	50.0	2.61	10.0	0.86
129.0	5.86	89.0	4.25	49.0	2.56	9.0	0.81
128.0	5.82	88.0	4.21	48.0	2.52	8.0	0.77
127.0	5.79	87.0	4.17	47.0	2.48	7.0	0.72
126.0	5.75	86.0	4.13	46.0	2.44	6.0	0.68
125.0	5.71	85.0	4.09	45.0	2.39	5.0	0.63
124.0	5.67	84.0	4.05	44.0	2.35		
123.0	5.63	83.0	4.01	43.0	2.31		
122.0	5.59	82.0	3.96	42.0	2.26		
121.0	5.55	81.0	3.92	41.0	2.22		
120.0	5.51	80.0	3.88	40.0	2.18		
119.0	5.47	79.0	3.84	39.0	2.13		
118.0	5.43	78.0	3.80	38.0	2.09		
117.0	5.39	77.0	3.75	37.0	2.05		
116.0	5.35	76.0	3.71	36.0	2.00		
115.0	5.31	75.0	3.67	35.0	1.96		
114.0	5.27	74.0	3.63	34.0	1.92		
113.0	5.23	73.0	3.59	33.0	1.87		
112.0	5.19	72.0	3.54	32.0	1.83		
111.0	5.15	71.0	3.50	31.0	1.78		