



166 Keystone Drive
 Montgomeryville, PA 18936
 TEL: 215-641-2700
 FAX: 215-619-0458

UNITS	SCCM AIR	DATE	March 19, 2002
TUBE NUMBER	600(E100B) TYPICAL	STD CONDITIONS	1 ATM & 70 DEG F
FLOAT NUMBER	OO2	GAS TEMPERATURE	70 DEG F
FLOAT MATERIAL	GLASS	PRESSURE IN TUBE	0 PSIG

SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE
150.0	55.78	110.0	32.50	70.0	15.35	30.0	5.00
149.0	55.32	109.0	31.94	69.0	15.04	29.0	4.78
148.0	54.84	108.0	31.39	68.0	14.74	28.0	4.56
147.0	54.34	107.0	30.84	67.0	14.44	27.0	4.35
146.0	53.84	106.0	30.30	66.0	14.15	26.0	4.13
145.0	53.32	105.0	29.77	65.0	13.86	25.0	3.93
144.0	52.78	104.0	29.24	64.0	13.57	24.0	3.73
143.0	52.24	103.0	28.72	63.0	13.28	23.0	3.53
142.0	51.69	102.0	28.21	62.0	13.00	22.0	3.34
141.0	51.13	101.0	27.70	61.0	12.72	21.0	3.15
140.0	50.56	100.0	27.20	60.0	12.45	20.0	2.97
139.0	49.98	99.0	26.71	59.0	12.18	19.0	2.80
138.0	49.39	98.0	26.23	58.0	11.90	18.0	2.63
137.0	48.80	97.0	25.75	57.0	11.64	17.0	2.47
136.0	48.21	96.0	25.28	56.0	11.37	16.0	2.32
135.0	47.60	95.0	24.81	55.0	11.11	15.0	2.18
134.0	47.00	94.0	24.36	54.0	10.85	14.0	2.05
133.0	46.39	93.0	23.91	53.0	10.59	13.0	1.94
132.0	45.77	92.0	23.46	52.0	10.33	12.0	1.83
131.0	45.16	91.0	23.03	51.0	10.07	11.0	1.73
130.0	44.54	90.0	22.60	50.0	9.82	10.0	1.65
129.0	43.92	89.0	22.18	49.0	9.56	9.0	1.58
128.0	43.30	88.0	21.76	48.0	9.31	8.0	1.53
127.0	42.68	87.0	21.36	47.0	9.06	7.0	1.49
126.0	42.06	86.0	20.95	46.0	8.81	6.0	1.47
125.0	41.44	85.0	20.56	45.0	8.56	5.0	1.47
124.0	40.82	84.0	20.17	44.0	8.31		
123.0	40.21	83.0	19.79	43.0	8.07		
122.0	39.59	82.0	19.42	42.0	7.82		
121.0	38.98	81.0	19.05	41.0	7.58		
120.0	38.37	80.0	18.68	40.0	7.34		
119.0	37.76	79.0	18.33	39.0	7.09		
118.0	37.16	78.0	17.97	38.0	6.85		
117.0	36.56	77.0	17.63	37.0	6.62		
116.0	35.97	76.0	17.29	36.0	6.38		
115.0	35.38	75.0	16.95	35.0	6.14		
114.0	34.79	74.0	16.62	34.0	5.91		
113.0	34.21	73.0	16.30	33.0	5.68		
112.0	33.63	72.0	15.98	32.0	5.45		
111.0	33.06	71.0	15.66	31.0	5.22		