



NITS : SCCM PROPANE : DATE : 10-27-1992  
UBE NUMBER : 601 (E200) : STD CONDITIONS : 1 ATMOS. & 70 DEG F  
ERIAL NUMBER : TYPICAL : GAS TEMPERATURE : 70 DEG F  
LOAT MATERIAL: GLASS : PRESSURE IN TUBE : 0 PSIG

| SCALE<br>READING | FLOW<br>RATE | SCALE<br>READING | FLOW<br>RATE | SCALE<br>READING | FLOW<br>RATE | SCALE<br>READING | FLOW<br>RATE |
|------------------|--------------|------------------|--------------|------------------|--------------|------------------|--------------|
| 50.0             | 76.1900      | 110.0            | 43.2696      | 70.0             | 20.5561      | 30.0             | 7.2083       |
| 49.0             | 75.2478      | 109.0            | 42.5767      | 69.0             | 20.1160      | 29.0             | 6.9741       |
| 48.0             | 74.3115      | 108.0            | 41.8903      | 68.0             | 19.6819      | 28.0             | 6.7441       |
| 47.0             | 73.3809      | 107.0            | 41.2105      | 67.0             | 19.2536      | 27.0             | 6.5181       |
| 46.0             | 72.4562      | 106.0            | 40.5370      | 66.0             | 18.8312      | 26.0             | 6.2961       |
| 45.0             | 71.5374      | 105.0            | 39.8701      | 65.0             | 18.4146      | 25.0             | 6.0781       |
| 44.0             | 70.6245      | 104.0            | 39.2097      | 64.0             | 18.0039      | 24.0             | 5.8639       |
| 43.0             | 69.7176      | 103.0            | 38.5558      | 63.0             | 17.5989      | 23.0             | 5.6535       |
| 42.0             | 68.8166      | 102.0            | 37.9083      | 62.0             | 17.1997      | 22.0             | 5.4469       |
| 41.0             | 67.9216      | 101.0            | 37.2673      | 61.0             | 16.8061      | 21.0             | 5.2439       |
| 40.0             | 67.0326      | 100.0            | 36.6328      | 60.0             | 16.4182      | 20.0             | 5.0445       |
| 39.0             | 66.1497      | 99.0             | 36.0048      | 59.0             | 16.0360      | 19.0             | 4.8486       |
| 38.0             | 65.2729      | 98.0             | 35.3832      | 58.0             | 15.6593      | 18.0             | 4.6562       |
| 37.0             | 64.4022      | 97.0             | 34.7681      | 57.0             | 15.2882      | 17.0             | 4.4672       |
| 36.0             | 63.5376      | 96.0             | 34.1594      | 56.0             | 14.9226      | 16.0             | 4.2815       |
| 35.0             | 62.6791      | 95.0             | 33.5572      | 55.0             | 14.5625      | 15.0             | 4.0991       |
| 34.0             | 61.8269      | 94.0             | 32.9614      | 54.0             | 14.2078      | 14.0             | 3.9199       |
| 33.0             | 60.9808      | 93.0             | 32.3721      | 53.0             | 13.8586      | 13.0             | 3.7437       |
| 32.0             | 60.1409      | 92.0             | 31.7891      | 52.0             | 13.5146      | 12.0             | 3.5706       |
| 31.0             | 59.3073      | 91.0             | 31.2126      | 51.0             | 13.1760      | 11.0             | 3.4005       |
| 30.0             | 58.4799      | 90.0             | 30.6424      | 50.0             | 12.8427      | 10.0             | 3.2333       |
| 29.0             | 57.6589      | 89.0             | 30.0786      | 49.0             | 12.5146      | 9.0              | 3.0688       |
| 28.0             | 56.8441      | 88.0             | 29.5212      | 48.0             | 12.1916      | 8.0              | 2.9071       |
| 27.0             | 56.0356      | 87.0             | 28.9702      | 47.0             | 11.8738      | 7.0              | 2.7481       |
| 26.0             | 55.2334      | 86.0             | 28.4255      | 46.0             | 11.5611      | 6.0              | 2.5916       |
| 25.0             | 54.4376      | 85.0             | 27.8871      | 45.0             | 11.2534      | 5.0              | 2.4376       |
| 24.0             | 53.6481      | 84.0             | 27.3550      | 44.0             | 10.9507      |                  |              |
| 23.0             | 52.8651      | 83.0             | 26.8291      | 43.0             | 10.6530      |                  |              |
| 22.0             | 52.0884      | 82.0             | 26.3096      | 42.0             | 10.3601      |                  |              |
| 21.0             | 51.3180      | 81.0             | 25.7963      | 41.0             | 10.0721      |                  |              |
| 20.0             | 50.5541      | 80.0             | 25.2892      | 40.0             | 9.7889       |                  |              |
| 19.0             | 49.7967      | 79.0             | 24.7884      | 39.0             | 9.5104       |                  |              |
| 18.0             | 49.0456      | 78.0             | 24.2937      | 38.0             | 9.2366       |                  |              |
| 17.0             | 48.3010      | 77.0             | 23.8052      | 37.0             | 8.9674       |                  |              |
| 16.0             | 47.5628      | 76.0             | 23.3228      | 36.0             | 8.7028       |                  |              |
| 15.0             | 46.8311      | 75.0             | 22.8465      | 35.0             | 8.4428       |                  |              |
| 14.0             | 46.1059      | 74.0             | 22.3764      | 34.0             | 8.1872       |                  |              |
| 13.0             | 45.3871      | 73.0             | 21.9123      | 33.0             | 7.9360       |                  |              |
| 12.0             | 44.6748      | 72.0             | 21.4542      | 32.0             | 7.6891       |                  |              |
| 11.0             | 43.9690      | 71.0             | 21.0022      | 31.0             | 7.4466       |                  |              |

ask...The Gas Professionals™  
For a complete library of rotameter calibration data, visit:  
http://www.mathesongas.com/gas-handling/rotameter-FAQ