

# Matheson GAS EQUIPMENT TECHNOLOGY GROUP



166 Keystone Drive, Montgomeryville, PA 18936 • Phone (215) 641-2700 • FAX (215) 641-2714

UNITS : SLPM NITROGEN TRIFLUORIDE  
 DATE : 08/27/96 (F)  
 TUBE NUMBER : 603 (E500) STD CONDITIONS : 1 ATMOS. & 70 DEG F  
 SERIAL NUMBER : TYPICAL GAS TEMPERATURE : 70 DEG F  
 FLOAT MATERIAL: GLASS PRESSURE IN TUBE : 0 PSIG

SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE
150.0	1.54	110.0	1.22	70.0	0.84	30.0	0.39
149.0	1.53	109.0	1.21	69.0	0.83	29.0	0.38
148.0	1.53	108.0	1.20	68.0	0.82	28.0	0.37
147.0	1.52	107.0	1.19	67.0	0.81	27.0	0.35
146.0	1.52	106.0	1.18	66.0	0.80	26.0	0.34
145.0	1.51	105.0	1.18	65.0	0.79	25.0	0.33
144.0	1.50	104.0	1.17	64.0	0.78	24.0	0.31
143.0	1.50	103.0	1.16	63.0	0.77	23.0	0.30
142.0	1.49	102.0	1.15	62.0	0.76	22.0	0.29
141.0	1.48	101.0	1.14	61.0	0.75	21.0	0.27
140.0	1.47	100.0	1.13	60.0	0.73	20.0	0.26
139.0	1.47	99.0	1.12	59.0	0.72	19.0	0.24
138.0	1.46	98.0	1.11	58.0	0.71	18.0	0.23
137.0	1.45	97.0	1.10	57.0	0.70	17.0	0.21
136.0	1.44	96.0	1.09	56.0	0.69	16.0	0.20
135.0	1.44	95.0	1.08	55.0	0.68	15.0	0.18
134.0	1.43	94.0	1.07	54.0	0.67	14.0	0.17
133.0	1.42	93.0	1.06	53.0	0.66	13.0	0.15
132.0	1.41	92.0	1.05	52.0	0.65	12.0	0.14
131.0	1.40	91.0	1.04	51.0	0.64	11.0	0.12
130.0	1.40	90.0	1.03	50.0	0.63	10.0	0.10
129.0	1.39	89.0	1.02	49.0	0.62	9.0	0.09
128.0	1.38	88.0	1.01	48.0	0.61	8.0	0.07
127.0	1.37	87.0	1.00	47.0	0.60	7.0	0.05
126.0	1.36	86.0	0.99	46.0	0.58	6.0	0.04
125.0	1.36	85.0	0.98	45.0	0.57	5.0	0.02
124.0	1.35	84.0	0.98	44.0	0.56		
123.0	1.34	83.0	0.97	43.0	0.55		
122.0	1.33	82.0	0.96	42.0	0.54		
121.0	1.32	81.0	0.95	41.0	0.53		
120.0	1.31	80.0	0.94	40.0	0.52		
119.0	1.30	79.0	0.93	39.0	0.50		
118.0	1.29	78.0	0.92	38.0	0.49		
117.0	1.29	77.0	0.91	37.0	0.48		
116.0	1.28	76.0	0.90	36.0	0.47		
115.0	1.27	75.0	0.89	35.0	0.46		
114.0	1.26	74.0	0.88	34.0	0.44		
113.0	1.25	73.0	0.87	33.0	0.43		
112.0	1.24	72.0	0.86	32.0	0.42		
111.0	1.23	71.0	0.85	31.0	0.41		