



166 Keystone Drive
 Montgomeryville, PA 18936
 TEL: 215-641-2700
 FAX: 215-619-0458

UNITS	SLPM AIR	DATE	September 17, 2001
TUBE NUMBER	604(E700) TYPICAL	STD CONDITIONS	1 ATM & 70 DEG F
FLOAT NUMBER	OO5	GAS TEMPERATURE	70 DEG F
FLOAT MATERIAL	STAINLESS STEEL	PRESSURE IN TUBE	0 PSIG

SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE
150.0	16.58	110.0	12.35	70.0	7.90	30.0	3.23
149.0	16.49	109.0	12.24	69.0	7.78	29.0	3.12
148.0	16.39	108.0	12.13	68.0	7.67	28.0	3.00
147.0	16.29	107.0	12.03	67.0	7.55	27.0	2.89
146.0	16.19	106.0	11.92	66.0	7.43	26.0	2.77
145.0	16.09	105.0	11.81	65.0	7.32	25.0	2.66
144.0	15.99	104.0	11.70	64.0	7.20	24.0	2.54
143.0	15.89	103.0	11.59	63.0	7.08	23.0	2.43
142.0	15.79	102.0	11.49	62.0	6.96	22.0	2.32
141.0	15.68	101.0	11.38	61.0	6.85	21.0	2.20
140.0	15.58	100.0	11.27	60.0	6.73	20.0	2.09
139.0	15.47	99.0	11.16	59.0	6.61	19.0	1.97
138.0	15.37	98.0	11.05	58.0	6.49	18.0	1.85
137.0	15.26	97.0	10.94	57.0	6.37	17.0	1.74
136.0	15.15	96.0	10.83	56.0	6.26	16.0	1.62
135.0	15.05	95.0	10.72	55.0	6.14	15.0	1.50
134.0	14.94	94.0	10.61	54.0	6.02	14.0	1.38
133.0	14.83	93.0	10.50	53.0	5.90	13.0	1.27
132.0	14.73	92.0	10.39	52.0	5.79	12.0	1.14
131.0	14.62	91.0	10.28	51.0	5.67	11.0	1.02
130.0	14.51	90.0	10.17	50.0	5.55	10.0	0.90
129.0	14.40	89.0	10.06	49.0	5.43	9.0	0.78
128.0	14.29	88.0	9.95	48.0	5.31	8.0	0.65
127.0	14.19	87.0	9.84	47.0	5.20	7.0	0.52
126.0	14.08	86.0	9.73	46.0	5.08	6.0	0.39
125.0	13.97	85.0	9.61	45.0	4.96	5.0	0.26
124.0	13.86	84.0	9.50	44.0	4.85		
123.0	13.75	83.0	9.39	43.0	4.73		
122.0	13.64	82.0	9.28	42.0	4.61		
121.0	13.54	81.0	9.16	41.0	4.50		
120.0	13.43	80.0	9.05	40.0	4.38		
119.0	13.32	79.0	8.94	39.0	4.26		
118.0	13.21	78.0	8.82	38.0	4.15		
117.0	13.10	77.0	8.71	37.0	4.03		
116.0	13.00	76.0	8.59	36.0	3.92		
115.0	12.89	75.0	8.48	35.0	3.80		
114.0	12.78	74.0	8.36	34.0	3.69		
113.0	12.67	73.0	8.25	33.0	3.57		
112.0	12.57	72.0	8.13	32.0	3.46		
111.0	12.46	71.0	8.02	31.0	3.34		